



TO SAFELY CHANGE A LINER:

1. Disconnect the patient tubing from the patient port and cap the port.
2. Carefully remove the tandem tube from the patient port (if it is a tandem setup).
3. Disconnect the red vacuum tube from the vacuum port and cap the port.
4. Confirm that all unused ports are securely capped off prior to turning off the suction.
5. Turn off the ON/OFF valve.
6. Carefully remove the liner from the outer canister.